

Technical Bulletin - Swabs

The LumiraDx SARS-CoV-2 & Flu A/B test is an automated rapid microfluidic immunofluorescence assay for use with the LumiraDx Platform, for near-patient testing, intended for the qualitative detection and differentiation of SARS-CoV-2, Influenza A and/or Influenza B viral antigens from nasal swab samples. Samples are collected from individuals suspected of respiratory viral infection consistent with COVID-19 by their healthcare provider.

This Technical Bulletin will aim to provide a comprehensive list of LumiraDx validated swabs for use with the SARS-CoV-2 & Flu A/B test.

Swab name	Supplier Product Code
Copan Nasal FLOQSwab [®] Regular with 30mm breakpoint	520CS01
Copan Nasal FLOQSwab [®] Regular with 80mm breakpoint	502CS01
Copan Nasal FLOQSwab [®] Regular with100mm breakpoint	519CS01
Copan Nasal FLOQSwab [®] Regular with no breakpoint	5U060S01
Copan FLOQswab [®] Regular Tip with 30mm breakpoint (in dry tube)	520C
Copan FLOQswab [®] Regular Tip with 80mm breakpoint (in dry tube)	552C
Copan FLOQswab [®] Regular Tip with 100mm breakpoint (in dry tube)	519C
Copan FLOQswab [®] Minitip with 80mm breakpoint	501CS01
Copan FLOQswab [®] Minitip with 100mm breakpoint	518CS01
Copan FLOQswab [®] Minitip with 80mm breakpoint (in dry tube)	518C
Copan FLOQswab [®] Minitip with 100mm breakpoint (in dry tube)	551C
Copan FLOQswab [®] Mini Flexible with 100mm breakpoint	516CS01
Copan FLOQSwab [®] Mini Flexible with no breakpoint	535C\$01
Copan FLOQSwab [®] Mini Flexible 100mm breakpoint (in dry tube)	516C
Copan FLOQSwab [®] Mini Flexible no break point (in dry tube)	535C
Puritan [®] HydraFlock Sterile Standard Flock Swab	25-3506-H
mwe medical wire DrySwab™ Rayon Swab	MW112
Gongdong Disposable Sterile Swab with 77mm breakpoint	G1029FT77

S-COM-ART-02176 R2 Rev Date 2024-05



Technical Bulletin - Swabs

This Technical Bulletin includes a complete list of swabs specific to a supplier and a product code whose performance was confirmed acceptable for use with the LumiraDx SARS-CoV-2 & Flu A/B test at the time of validation testing. LumiraDx cannot guarantee that swab suppliers will not change raw materials or other swab specifications, which could affect performance. Please review the swab supplier's limitations and intended use before sample collection. LumiraDx cannot guarantee the availability of the swabs listed.

Swabs from alternative suppliers have not been validated for use with the LumiraDx SARS-CoV-2 & Flu A/B test. The Test may not perform as expected if using swabs from alternative suppliers.

Furthermore, the above suppliers may offer similar swabs under a different product code. These swabs have not been validated for use with the LumiraDx SARS-CoV-2 & Flu A/B test. The test may not perform as expected if using swabs other than the exact product codes listed above.

For product enquiries and technical support please contact LumiraDx Customer Services.

Email: customerservices@lumiradx.com

Phone: +44 (0)1172841222

www.lumiradx.com

Copyright © 2024 LumiraDx UK LTD. All rights reserved worldwide. LumiraDx and Flame logo are trademarks of LumiraDx International LTD. Full details of these and other registrations of LumiraDx can be found at lumiradx.com/IP. All other trademarks are the property of their respective owners. Content should be used for the use of the LumiraDx products only and in line with instructions provided. You may not, except with our express written permission, distribute or commercially exploit the content. Nor may you transmit it or store it in any other form of electronic retrieval system other than for the purpose of use of the LumiraDx Instrument or LumiraDx Test Strips. Information provided is subject to change without notice.

Product is not available in all countries and regions. Please check with your local LumiraDx sales representative or distributor for availability in specific markets.

Manufactured by:

LumiraDx UK Ltd Dumyat Business Park Alloa FK10 2PB UK

Registration Number: 09206123

Authorized representative in the European Union: LumiraDx AB, Västra Vägen 5A, 16961 Solna, Sweden